UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/544,341	04/06/2000	Kevin A. Mansmann, M.D.	KM-M1	8036
29369 PATRICK D. I	7590 02/08/2001 CELLY	EXAMINER		
11939 MANCHESTER #403			ISABELLA, DAVID J	
ST. LOUIS, MO 63131		•	ART UNIT	PAPER NUMBER
			3738	
		·		
SHORTENED STATUTOR	Y PERIOD OF RESPONSE	MAIL DATE	DELIVERY MODE	
3 MONTHS		02/08/2007	PAPER	

Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

	Application No.	Applicant(s)				
Office Action Comments	09/544,341	MANSMANN, M.D., KEVIN A.				
Office Action Summary	Examiner	Art Unit				
	DAVID J. ISABELLA	3738				
The MAILING DATE of this communication app Period for Reply	ears on the cover sheet with the c	orrespondence address				
A SHORTENED STATUTORY PERIOD FOR REPLY WHICHEVER IS LONGER, FROM THE MAILING DA - Extensions of time may be available under the provisions of 37 CFR 1.13 after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period w - Failure to reply within the set or extended period for reply will, by statute, Any reply received by the Office later than three months after the mailing earned patent term adjustment. See 37 CFR 1.704(b).	ATE OF THIS COMMUNICATION 16(a). In no event, however, may a reply be time rill apply and will expire SIX (6) MONTHS from a cause the application to become ABANDONEI	l. hely filed the mailing date of this communication.				
Status						
1) Responsive to communication(s) filed on 20 No	Responsive to communication(s) filed on 20 November 2006.					
3) Since this application is in condition for allowar	3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is					
closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213.						
Disposition of Claims						
4)⊠ Claim(s) <u>24-32</u> is/are pending in the application.						
4a) Of the above claim(s) <u>29-32</u> is/are withdrawn from consideration.						
5) Claim(s) is/are allowed.						
6)⊠ Claim(s) <u>24-28</u> is/are rejected.						
7) Claim(s) is/are objected to.						
8) Claim(s) are subject to restriction and/or	election requirement.					
Application Papers						
9) The specification is objected to by the Examiner	•					
10) The drawing(s) filed on is/are: a) accepted or b) objected to by the Examiner.						
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).						
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).						
11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.						
Priority under 35 U.S.C. § 119						
12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).						
a) All b) Some * c) None of:						
1. Certified copies of the priority documents have been received.						
2. Certified copies of the priority documents have been received in Application No						
3. Copies of the certified copies of the priority documents have been received in this National Stage						
application from the International Bureau (PCT Rule 17.2(a)).						
* See the attached detailed Office action for a list of the certified copies not received.						
 Attachmant(s)	•	*				
Attachment(s)						
1) Notice of References Cited (PTO-892) 2) Notice of Draftsperson's Patent Drawing Review (PTO-948)	4) Interview Summary Paper No(s)/Mail Da					
3) Information Disclosure Statement(s) (PTO/SB/08)	5) Notice of Informal Pa					
Paper No(s)/Mail Date	6) Other:					

Status of the Claims

Applicant's election without traverse of claims 24-28 in the reply filed on 11/20/2006 is acknowledged.

In reply to the restriction requirement mailed on October 20, 2006, Applicant elected the claims of Group I and species relating to cartilage repair (Claims 24-28).

Claim Rejections - 35 USC § 112

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claim 28 is rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

The limitation of "steps comprising surface treatment of a thick permeable material in a manner that creates a modified surface layer" fails to positively set forth structure that defines the particular product. As worded the product by process steps is ambiguous, in that, the "metes and bounds" of the scope of the claim are not readily discernable to one with ordinary skills in the art. Therefor, the claim, in so far as definite, will be treated as a product claim where in the segment has a surface layer.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

Application/Control Number: 09/544,341

Art Unit: 3738

(a) the invention was known or used by others in this country, or patented or described in a printed publication in this or a foreign country, before the invention thereof by the applicant for a patent.

Claims 24-28 are rejected under 35 U.S.C. 102(a) as being anticipated by Meenen et al [6319712].

Meenen et al discloses a synthetic membrane segment for surgically treating damaged cartilage in a mammalian joint, comprising a membrane segment made of a synthetic polymer, wherein said membrane segment is hydrophilic and permeable to water but is not permeable to surface-active phospholipids or hyaluronate

Molecules (see column 4, lines 30+), and wherein said membrane segment has at least one surface suited for use as an articulating surface in a joint, wherein said articulating surface is designed to interact with water, surface-active phospholipid molecules, and hyaluronate molecules in natural synovial fluid, after surgical implantation into a joint, in a manner that sustains lubrication of said articulating surface by said water, surface-active phospholipid molecules, and hyaluronate molecules (column 6, lines 20+).

Claim 25, the membrane is affixed to a resorbable scaffold that will support cell growth, inlcuind chondrocyte and cartilage regeneration (column 1, lines 45+).

Claim 26, the membrane can be trimmed prior to application to the damaged surface area on the cartilage surface.

Claim 27, see anchoring sutures in column 5, lines 45+.

Claim 28, in so far as the final product is defined by the method step, the claimed product fails to distinguish over the same as disclosed by Meenen et al. Meenen et al teaches a modified surface layer.

Art Unit: 3738

Response to Arguments

Applicant's arguments with respect to claims 24-28 have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to DAVID J. ISABELLA whose telephone number is 571-272-4749. The examiner can normally be reached on MONDAY-FRIDAY.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, CORRINE MCDERMOTT can be reached on 571-272-4754. The fax phone

Art Unit: 3738

number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) er-571-272-1000.

DAVID J ISABELLA Primary Examiner Art Unit 3738

DJI 2/1/2007